

**MODEL: R4KPB11 — PUSH FOR HELP STATION**



R4KPB11



**FEATURES**

- Call-Assurance LED
- Extra large, back-lit pushbutton for easy activation
- Cancel touchpoint
- Continuous supervision for station presence
- Unique mounting plate design accommodates a variety of single and dual gang backboxes
- Plug-in connectors for easy service

**SPECIFICATIONS**

**Power Required:** 15.5V DC @ 4mA

**Weight:** 0.25 lb. (0.11 kg)

**Size:** W: 4.21" (10.69 cm)

H: 5.46" (13.87 cm)

D: 1.17" (2.97 cm)

**Housing & Finish:** High-impact molded ABS94VO;  
light gray faceplate

**Controls:** One (1) Call touchpoint

One (1) Cancel button

One (1) Call-Assurance indicator

**Terminations:** Plug-on connectors

**Backbox Options:** Raco 561, Steel City 58371 3/4R or any other single gang backbox. Mounting plate allows for North American, South American, and European style single and dual gang backboxes

**Certification:** UL/C-UL 1069; CE.

**DESCRIPTION**

The R4KPB11 is an emergency pushbutton station that provides the ability to place a unique call-in priority to the assigned console(s). The presence of a Cancel button allows for cancellation at the station. This type of station is normally located at a patient room head wall for placing a call for help, staff emergency and/or code call.

The Push For Help station is connected to its associated corridor light or domeless room controller and is continually supervised for station presence. Through programming, a "Trouble" message can annunciate at the associated console(s) when the station is removed from the system.

**ASSOCIATED EQUIPMENT**

CLA222 / CLA244 — Audio Corridor Lights

CLV122 / CLV144 — Visual Corridor Lights

DCA200 / DCA216 — Audio Domeless Room Controllers

DCV100 / DCV116 — Visual Domeless Room Controllers

*Architect and Engineer (A&E) Specifications available online at: [customerconnection.rauland.com](http://customerconnection.rauland.com)  
Specifications subject to change without notice*

©Copyright 2014 Rauland-Borg Corporation Printed in USA Rev 04/2014