

MODEL: 353010 - DUAL PATIENT STATION



FEATURES

- Multiple call-in priorities available
- Two Tilt/Release DIN receptacles for Pillow Speakers with patient entertainment controls and handset intercom with entertainment muting during nurse call
- Tilt/Release DIN receptacle allows for ease of disconnecting Pillow Speaker from the patient station
- Two auxiliary device receptacles with separate call-in priority settings
- No dummy plugs required
- Interfaces with specialty beds via separate receptacle at rear of station
- LEDs to indicate specialty bed(s) are connected
- Two Call-Assurance LEDs to indicate calls placed from the pillow speakers
- Two Call-Assurance LEDs to indicate calls placed from the ¼" receptacles
- Outputs for lighting control options direct to low voltage lighting controllers
- Monitor LED indicates audio status
- Cancel button for localized call cancellation
- Continuous supervision of station for loss of data and/or power faults
- Plug-in connectors for easy service
- Bio-seal option to cover buttons and allow for easy disinfecting of the station

SPECIFICATIONS

POWER REQUIRED: 220mA at 18V (full load all connections)

WEIGHT: 0.70 lb. (0.317 kg)

SIZE: W: 6.39" (16.2 cm)

H: 4.25" (10.8 cm)

D: 2.00" (5.1) in back box; 0.50" (1.3 cm) off-wall

HOUSING AND FINISH: High-impact molded Cyclopol 2950 UL 94-5VB; light grey faceplate

CONTROLS: Two (2) DIN receptacles, Two (2) ¼" receptacles, One

(1) Cancel button, One (1) Monitor LED, Two (2) Call-assurance LED indicators, and Two (2) Special Bed LED indicators

TERMINATIONS: Plug-on connector

BACKBOX REQUIREMENT: 3 Device, Non-gangable, 3-1/2" Deep (i.e. Raco 697, Steel City GW-335C, Steel City H3BD-3/4 with 3GC plater ring, or UL recognized equivalent metal electrical box)

CERTIFICATION: UL/C-UL 1069, FCC Part 15 Class B, IEC60601-1 and IEC60601-1-2, CE Mark (LVD) (all Pending)

DESCRIPTION

The Dual Patient Station allows both patients and staff to place different call-ins, depending on their requirements, to the assigned nurse console(s), staff terminal(s), along with other annunciation devices such as wireless phones. The patient station interfaces with entertainment handsets/Pillow Speakers and specialty beds to initiate calls, control entertainment, control lighting, and provide nurse call audio and entertainment audio through the respective speakers. When used with the Feature Bed Control Module, the audio capabilities are transferred to the bed rail with entertainment muting during nurse communication.

The two beds share Nurse Call activity but with separate TV and lighting control through pillow speakers.

The patient station is connected to its associated corridor light or domeless room controller and is continually supervised for both power and data. Through programming, a message can annunciate at the associated console(s) and / or the network based remote status application if a problem occurs.

Rauland-Borg Corporation

Architects and Engineers Specifications available on disc. Specifications subject to change without notice.

USA • 800-752-7725 • Fax 800-217-0977
CANADA • 905-607-2335 • Fax 905-607-3554
www.rauland.com

MODEL #353010
05/2008