

MODEL: 354001 - PULL CORD STATION



FEATURES

- Plastic pull cord for easy activation
- Smooth, non-porous PVC coating on the cord provides an easily cleaned surface
- Call-Assurance LED
- Cancel button
- Water-resistant
- Large, easy to grab bell
- Continuous supervision of station
- Plug-in connectors for easy service

SPECIFICATIONS

POWER REQUIRED: 4.74mA @ 18VDC

WEIGHT: 0.22 lbs (0.10 kg)

SIZE: W: 3.13" (8.1 cm)

H: 4.25" (10.8 cm)

D: 0.63" (1.6 cm) in back box: 0.50" (1.3 cm) off-wall

HOUSING & FINISH: High-impact molded Cyclopol 2950 UL 94-5VB; light grey faceplate

CONTROLS: One (1) pull cord for call actuation, One (1) Call-assurance LED, One (1) Cancel button

TERMINATIONS: Plug-on connector

BACKBOX REQUIREMENT: 1 device, Non-gangable, 3 1/2" deep, (i.e. Raco 561, Steel City 58371, 3/4 R, or UL recognized equivalent metal backbox)

CERTIFICATION: UL/C-UL 1069, FCC Part 15 Class B, IEC60601-1 and IEC60601-1-2, CE Mark (LVD) (all pending)

DESCRIPTION

The Pull Cord station is an emergency pull cord station that provides the ability to place a unique call-in priority from a single station by pulling on its plastic cord. The Pull-cord station is typically located in bathrooms near the toilet and in the shower. The membrane Cancel button allows for cancellation of the call at the station. Unit may utilize remote Cancel button(s) for ease of nurse call cancel when pull cord is mounted in an inaccessible location (e.g. ceiling). To make the station water-resistant, the station perimeter is caulked on three sides with

sealant. It also has water diversion channels that open along the bottom of the station for drainage. The Pull-cord station is connected to its associated corridor light or Domeless room controller and is continually supervised.

Rauland-Borg Corporation

Architects and Engineers Specifications available on disc. Specifications subject to change without notice.

USA • 800-752-7725 • Fax 847-679-4106
CANADA • 905-607-2335 • Fax 905-607-3554
www.rauland.com

MODEL #354001
05/2008