



Ref. Certif. No.

DK-70900-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

## CB TEST CERTIFICATE

Product

Medical, Dental Power Supply

Name and address of the applicant

V GULDMANN A/S  
GRAHAM BELLS VEJ 21-23A  
8200 AARHUS N DENMARK

Name and address of the manufacturer

V GULDMANN A/S  
GRAHAM BELLS VEJ 21-23A  
8200 AARHUS N DENMARK

Name and address of the factory

Note: When more than one factory, please report on page 2

FOSHAN NORATEL ELECTRIC CO LTD  
JUNYE RD, ZONE C OF SHISHAN INDUSTRIAL PARK,  
NANHAI DISTRICT, FOSHAN  
GUANDONG 522825 CHINA☐ Additional Information on page 2

Ratings and principal characteristics

DK-13991: 100-115Vac, 50-60Hz, 1A or 230Vac, 50-60Hz, 0.5A  
or 220Vac, 60Hz, 0.5A  
DK-14106: 100-115Vac, 50-60Hz, 1A or 230Vac, 50-60Hz, 0.5A  
or 220 Vac, 60Hz 0.5A

Trademark (if any)

None

Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

DK-13991. DK-14106

Additional information (if necessary may also be  
reported on page 2)

National Differences specified in the CB Test Report.

☐ Additional Information on page 2A sample of the product was tested and found  
to be in conformity withIEC 60601-1(ed.3), IEC 60601-1(ed.3);am1, IEC 60601-1-  
6(ed.3), IEC 60601-1-6(ed.3);am1As shown in the Test Report Ref. No. which forms part  
of this Certificate

E351786-D1000-1/A0/C0-ULCB-ULCB issued on 2018-02-21

This CB Test Certificate is issued by the National Certification Body



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

Date: 2018-02-26

Signature:

Jan-Erik Storgaard

For full legal entity names see [www.ul.com/nbnames](http://www.ul.com/nbnames)