



CUSTOMER		SYSTEM	
Site name:	<u>Clínica Santa María</u>	Serial Number	<u>F0023</u>
Address:	<u>500 Sta María Santiago - Chile</u>	Installation date	<u>30/11/2015</u>
		Software version	<u>F3</u>
Contact on site:	<u>CRISTIAN LOPEZ</u>	Probe(s) S/N	<u>UV186 + UV186(s)</u> ☐ NA
		HVPS shocks	☒ NA

INTERVENTION DETAILS						
Work carried out:	☐ Warranty	☒ Service contract	☐ Loan	☐ RPP	☐ Billable	☐ _____
	☐ Installation	☐ Preventive	☐ Curative	☐ Application	☐ Upgrade	☐ _____
	* We installed probe F0 UV186 - IR probe UV147 * We went through the checklist: everything is OK. * Simulations done with the two F0 UV probes: OK. * FCA-209 - Focal One Re-treatment protocol update → OK					
Comments:	* One treatment followed on Wednesday 5th July with Dr Velasco The equipment is working fine					
Resolution code:	_____					

Customer observations:	_____
------------------------	-------

EQUIPMENTS USED					
Type of tool	Serial Number	Validity date	Type of tool	Serial Number	Validity date
Multimeter	<u>SAV290</u>	<u>8/01/24</u>	Antistatic protection	<u>SAV1359</u>	<u>06/09/2023</u>
<u>See checklist.</u>	_____	_____	_____	_____	_____

Field Service Engineer	Customer	Service Manager
Name: <u>Jorge F. Peña C.</u>	Name: <u>CRISTIAN LOPEZ</u>	Name: _____
Signature:	Signature:	Signature: _____
Start date: <u>5/07/2023</u>	End date: <u>13/07/2023</u>	Duration: <u>Revisión programada 3D (26h)</u>

Once completed, please send it back to ccc@edap-tms.com.