

# Performance Test Litho 35

Model of the system:

Serial Number:

## Optical Test (with user autoadjust on)

USE ONLY SERVICE FIBER	Hz / J	Power on the system display	Power read on the power meter	Lamp (us)
Lithotripsy mode				
365	10Hz / 3.5J			
365	10Hz / 1.2J			
365	12Hz / 1.0J			
200	10Hz / 1.2J			
200	12Hz / 1.0J			

If the results are in max 5% different you can do the next step

## Long run test (10 minutes lasering in continuous Lithotripsy mode)

200	12Hz / 1.0J	Temperature of the fiber jack	cold <input type="checkbox"/>	warm <input type="checkbox"/>	hot <input type="checkbox"/>
Tested By:		Surname:	Signature:		Date: