

Performance Test Cyber Ho 100						
Model of the system:						
Serial Number:						
Fiber Policy:	A		B		C	D
Number of the lamps shots:	L1		L2		L3	
P/N + S/N charger:	P/N		S/N			
P/N + S/N dischargers lamp 1:	P/N		S/N			
P/N + S/N dischargers lamp 2:	P/N		S/N			
P/N + S/N dischargers lamp 3:	P/N		S/N			
Optical Test (with user autoadjust on)						
Fiber	Settings Hz / J	Power on the system display	Power read on the power meter	Lamp 1 (us)	Lamp 2 (us)	Lamp 3 (us)
Short Pulse (Master Pulse 0)						
365	30Hz / 3.5J					
365	35Hz / 0.9J					
365	80Hz / 0.2J					
200	35Hz / 0.9J					
200	60Hz / 0.5J					
Medium Pulse (Master Pulse 3)						
365	30Hz / 3.4J					
365	35Hz / 0.9J					
200	35Hz / 0.9J					
200	60Hz / 0.5J					
Long Pulse (Master Pulse 6)						
365	30Hz / 3.3J					
365	35Hz / 0.9J					
200	35Hz / 0.9J					
200	60Hz / 0.5J					
Long run test (10 minutes lasering in continuous short pulse)						
200	35Hz / 0.9J	Temperature of the fiber jack	cold ☐	warm ☐	hot ☐	
Tested By:		Surname:	Signature:			Date: